

Short analytical evaluation of the new Beckman Coulter DxC 500i clinical analyzer

H. Baum, A. Sirch, C. Hartmann

Inst. für Laboratoriumsmedizin und Transfusionsmedizin, Standort Bietigheim, RKH Regionale Kliniken Holding und Services GmbH, Ludwigsburg, Germany

INTRODUCTION

The new DxC 500i clinical analyzer from Beckman Coulter combines two analytical units for general chemistry including ISE and for immunochemistry analysis. In this study we tested the analytical performance of this new consolidated system.

METHODS

Imprecision was evaluated using control material for 35 applications representing the majority of the assay menu. All parameters were measured over a period of 20 days twice a day. For selected parameters samples of external quality schemes were measured in duplicate twice a day for three days.

Comparison studies were done using samples from daily routine runs on a Cobas Pro system (Roche).

Statistics were done using Excel

RESULTS

Imprecision of the analytes of the clinical chemistry module as well as the ISE is shown in table and graph 1a and for the immuno module it is shown in table and graph 1b, respectively.

Comparison studies showed a good comparability with the cobas pro (table 2).

Results of the external QC-schemes are shown in table 3.

CONCLUSIONS

The new Beckman Coulter DxC 500i clinical analyzer integrates general chemistry and immunochemistry assays on a single platform for small or mid-size laboratories and shows accurate and precise results including a good method comparability.



RKH Gesundheit

RKH Regionale Kliniken Holding und Services GmbH

Table 3. Results of the external QC samples

RV CM4/23	sample	median	range	DxC500i mean	D %
CK-MB (ng/dL)	A	44.78	42.5 - 44.78	45.47	1.54
	B	67.1	64.95 - 72.1	70.25	4.69
RV HM4/23	sample	median	range	DxC500i mean	D %
	A	106	86.4 - 385.3	116.65	10.05
Ferritin (ng/mL)	A	140.1	106 - 423.3	148.35	5.89
	B	18.01	15.63 - 20.95	17.14	-4.85
fT3 (pg/mL)	A	3.15	2.76 - 32.61	3.06	-3.02
	B	1.15	0.83 - 12.678	1.14	-0.72
fT4 (ng/dL)	A	8.36	7.1 - 9.67	8.3	0.4
	B	6.72	6.72 - 34.73	6.71	-0.11
TSH (mU/mL)	A	24.32	9.43 - 27.5	23.34	-4.01
	B	2.32	6.72 - 34.73	6.71	-0.11
RV KS8/23	sample	median	range	DxC500i mean	D %
	A	106	102 - 109.3	104.47	-1.45
Cl (mmol/L)	A	98	96 - 101	97.3	-0.71
	B	3.82	3.71 - 4	3.83	0.26
K (mmol/L)	A	3.4	3.3 - 3.53	3.4	0.1
	B	123	119 - 127	122.57	-0.35
Na (mmol/L)	A	114	111 - 117	114.38	0.34
	B	3.52	3.26 - 3.7	3.61	2.3
Albumin (g/L)	A	3.09	2.81 - 3.5	3.18	3.02
	B	67	61 - 73.8	68.67	2.49
ALT (U/L)	A	43.2	40.6 - 49	43.67	1.08
	B	311.4	273 - 340	274	-12.01
ALP (U/L)	A	99	86 - 110	87.33	-11.78
	B	79	71 - 84	78	-1.27
Amylase (U/L)	A	49	44 - 52	48.5	-1.02
	B	209	183 - 228	197.67	-5.42
AST (U/L)	A	61	56 - 67	59.33	-2.73
	B	2.56	2.41 - 2.66	2.6	1.56
Ca (mmol/L)	A	178	17 - 184	182	2.06
	B	135	124 - 139	139.88	3.62
Chol (mg/dL)	A	116	108 - 133	119.78	3.26
	B	115	90 - 131	106.9	-7.04
C (K/L)	A	28.2	17.9 - 32.4	25.07	-11
	B	0.47	0.38 - 0.6	0.48	1.5
Crea (mg/dL)	A	2.12	1.89 - 2.38	2.24	5.42
	B	103	86 - 137	105.52	2.44
Iron (μg/dL)	A	94.92	84 - 140	98.53	3.81
	B	360.8	335 - 395	357.52	-0.91
GGT (U/L)	A	47	44 - 52	46.67	-0.71
	B	125.35	118 - 137	130.82	4.36
Glu (mg/dL)	A	68.63	65 - 75	71.88	4.73
	B	146	128 - 157	145.17	-0.57
LDH (U/L)	A	127.5	110 - 137	125.17	-1.83
	B	0.89	0.82 - 1.01	0.89	0
Mg (mmol/L)	A	0.79	0.72 - 1.02	0.81	2.95
	B	4.49	4.4 - 4.61	4.59	2.19
Phos, inorg (mg/dL)	A	2.57	2.51 - 2.69	2.61	1.43
	B	0.76	0.71 - 0.8	0.78	2.63
Tbil (mg/dL)	A	1.06	1.04 - 1.1	1.09	2.73
	B	5.58	5 - 5.8	5.62	0.66
TP (g/dL)	A	4.83	4.3 - 5.04	4.89	1.01
	B	87.85	Sep - 95	88.85	1.14
Tr (mg/dL)	A	62	58 - 68	63.58	2.55
	B	31.95	3.3 - 34.84	32.22	0.83
Urea (mg/dL)	A	18.65	17 - 21.62	19.02	1.97
	B	5.1	5.1 - 5.26	5.24	2.68
Uric (mg/dL)	A	4.24	4.2 - 4.34	4.32	2.05
	B	13	0.28 - 14.85	3.39	0.9139

Graph 1a Imprecision of ISE and the general chemistry assays

■ Level 1 CV % ■ Level 2 CV %




Table 1a Imprecision of ISE and the general chemistry assays

Analyte (unit)	Level 1		Level 2	
	Mean value	CV %	Mean value	CV %
Cl (mmol/L)	90.09	1.45	114.72	1.2
K (mmol/L)	4.06	1.31	6.81	1.12
Na (mmol/L)	121.93	1.43	153.94	1.11
Albumin (g/L)	24.14	0.9	45.55	0.82
ALP (U/L)	114.12	2.96	473.55	2.42
ALT (U/L)	40.66	1.85	125.65	1.48
Am yla se (U/L)	94.96	1.34	253.34	0.91
AST (U/L)	50.52	1.73	141.02	1.08
Ca (mmol/L)	2.39	1.47	3.21	1.78
Chol (mg/dL)	150.31	1.42	296.59	1.08
ck (U/L)	171.43	1.06	395.95	1.16
Crea (mg/dL)	1.26	1.22	5.93	0.98
CRP (mg/L)	12.31	0.84	83.39	0.93
GGT (U/L)	58.07	1.11	163.01	1.15
Glu (mg/dL)	97.42	1.2	237.39	0.79
HbA1c (mmol/mol)	39.37	1.72	69.82	1.25
HD L-C (mg/dL)	36.31	2.46	68.5	3.22
Iron (μg/dL)	69.51	2.62	212.8	2.43
LDH (U/L)	150.48	3.26	566.14	2.44
Mg (mmol/L)	0.99	2.61	1.62	2.49
Phos, inorg (mg/dL)	7.14	2.34	11.42	2.11
Tbil (mg/dL)	1.55	1.87	6.44	1.61
TP (g/dL)	3.82	2.28	7.53	2.26
Tr (mg/dL)	143.03	2.5	339.38	2.81
Urea (mg/dL)	37.34	1.84	169.03	1.8
Uric (mg/dL)	6.09	0.82	9.9	0.77

Graph 1b Imprecision of the immunological assays

■ Level 1 CV % ■ Level 2 CV % ■ Level 3 CV %




Table 1b Imprecision of the immunological assays

Analyte (unit)	Level 1		Level 2		Level 3	
	Mean value	CV %	Mean value	CV %	Mean value	CV %
CK-MB (μg/L)	4.38	4.67	55.83	3.56	17.09	5.27
Ferritin (μg/L)	46.94	4.22	399.59	6.09	148.62	4.39
fT4 (ng/dL)	1	4.25	2.03	2.45	3.35	3.28
fT3 (pg/mL)	3.04	4.51	5.97	4.13	10.48	4.15
HCG (mIU/mL)	3.2	3.44	272.11	4	104.14	6.67
IL-6 (pg/mL)	N/A	N/A	48.41	4.49	354.48	3.26
PCT (ng/mL)	N/A	N/A	1.15	4.4	11.13	5.12
TP (g/dL)	18.96	7.49	75.2	7.44	200.18	8.69
Tr (mg/dL)	3.17	3.72	10.52	3.62	29.8	5.04